Silicone rubber for Medical market

FUJIKURA COMPOSITES

Safe and Clean Silicone rubber molded into a specified shape for medical use

Biocompatibility Test	
Test	Result
Cytotoxicity test (ISO10993-5)	PASS
USP Class Ⅵ	PASS

Material	
Material	Content rate
Genetically modified organisms	NONE
of plant origin	NONE
of animal origin	NONE
Natural rubber/Latex	NONE

Additives and Impurities	
Material	Content rate
Phthalates	NONE
Residual solvent Class 1 - 3	NONE

Production

Integrated system of production (LIM)
of Silicone rubber for medical use



Feature of LIM

- Lower the risk of foreign matter mixed
- Precision and complex shape parts available
- Thin film (0.1 mm or less) molding available
- Continuous molding by automation (No-burr、Runnerless)
- Fast curing speed and shortened molding time

※2) Mirable Molding available by our partner.

< Contact >

FUJIKURA COMPOSITES
GOTANDA JP BUILDING 4F
8-4-13 Nishigotanda,Shinagawa-ku, Tokyo, Japan 141-0031
TEL. +81-3-5747-9223
FAX. +81-3-5747-9217

Web: https://www.fujikura-control.com/english
Founded: 1901
Sales: $281 million US dollar (2023 March, consolidated)
ISO certified: 9001, 14001, 13485
E-mail: seigyo.toiawase.en@fc.fujikura.co.jp